

INDIAN INSTITUTE OF TECHNOLOGY PATNA

(Store & Purchase Section)

Corrigendum-I

Date: 22.02.2022

Subject: Notice Inviting Tender (NIT) for supply of **“Pyrolysis Reactor”** to IIT Patna.

Tender Reference No.: **IITP/S&P/EPR/1/CE-407/2021-22, Dated: 02.02.2022.**

Tender ID: **2022_IITP_671658_1.**

I. In Technical Specification (Annexure-I), Sl. No. 3 is to be read as:

3. Temperature Control:

A Programmable, multi-level temp. controller will allow precise temp. profiles.

Rates upto 5 C / min will be possible.

Phase angle analog power control will ensure smooth power delivery and long Life of heater elements.

II. The last date of bid submission is being extended as per following schedule:

Last date and time for submission of quotation to be read as 01.03.2022 (till 03:00 pm)

Opening date and time of quotation to be read as 02.03.2022 (03:30 pm)

Other terms and conditions remain unchanged.